



MODERNWATER

Certificate of Analysis

Description: Optidose™ 96 Test

Part Number: 7000350

Lot Number: 12C4027

Expiration Date: August 2013

Storage Condition: Room Temperature

Quality Control Testing

| Test Specification (Results based on bottom line) | Results | Pass/Fail |
|---|---------|-----------|
| 2 ppm standard diluted to 2 ppm line: Equal to or Darker than Top Line | Equal | Pass |
| 2 ppm standard diluted to 5 ppm line: Darker than Top Line | Darker | Pass |
| 2 ppm standard diluted to 10 ppm line: Darker than Top Line | Darker | Pass |
| 5 ppm standard diluted to 2 ppm line: Lighter than Top Line | Lighter | Pass |
| 5 ppm standard diluted to 5 ppm line: Equal to or Lighter than Top Line | Equal | Pass |
| 5 ppm standard diluted to 10 ppm line: Darker than Top Line | Darker | Pass |
| 10 ppm standard diluted to 2 ppm line: Lighter than Top Line | Lighter | Pass |
| 10 ppm standard diluted to 5 ppm line: Lighter than Top Line | Lighter | Pass |
| 10 ppm standard diluted to 10 ppm line: Equal to or Lighter than Top Line | Equal | Pass |
| 7 ppm standard diluted to 10 ppm line: Darker than Top Line | Darker | Pass |

The values above are the mean of all Quality Control results obtained for this lot.

Reviewed by:

Sejal Patel
3/9/12

Modern Water Inc., 15 Read's Way, Suite 100, New Castle, DE 19720 USA

Tel: 1-302-669-6900; 1-855-637-6426; www.modernwater.com